

Instructions For Use CVK-1-IFU		
Rev. Date: July 31, 2008	Revision: 1	Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Fax (435) 755-0015 - www.scytek.com

Calcium Stain Kit (Modified Von Kossa)

Description: The Calcium Stain Kit (Modified Von Kossa) is intended for use in the histological visualization of calcium deposits in paraffin or frozen sections.

Calcium in mass deposits:	Black
Calcium in dispersed deposits:	Gray
Nuclei:	Red
Cytoplasm:	Light Pink

Uses/Limitations: For In-Vitro Diagnostic use only.
 Histological applications.
 Do not use past expiration date.
 Use caution when handling these reagents.

Control Tissue: Any paraffin embedded tissue that contains calcium deposits.

Availability/Contents:

<u>Item #</u>	<u>Kit Contents</u>	<u>Volume</u>	<u>Storage</u>
SNV125	Silver Nitrate Solution (5%)	125 ml	2-8° Centigrade
STB125	Sodium Thiosulfate Solution (5%)	125 ml	Room Temperature
NFS125	Nuclear Fast Red Solution	125 ml	Room Temperature


Precautions: Avoid contact with skin and eyes.
 Harmful if swallowed.
 Follow all Federal, State, and local regulations regarding disposal.
 Use in chemical fume hood whenever possible.

Procedure (Standard):


1. Deparaffinize sections if necessary and hydrate to distilled water.
2. Incubate slide in Silver Nitrate Solution (5%) for 30-60 minutes with exposure to ultraviolet light or 100 watt incandescent.
3. Rinse in several changes of distilled water.
4. Incubate slide in Sodium Thiosulfate Solution (5%) for 2-3 minutes.
5. Rinse for 2 minutes in running tap water followed by 2 changes of distilled water.
6. Incubate slide in Nuclear Fast Red Solution for 5 minutes.
7. Rinse for 2 minutes in running tap water followed by 2 changes of distilled water.
8. Dehydrate very quickly in 3 changes of fresh Absolute Alcohol.
9. Clear, and mount in synthetic resin.

References:

1. Sheenan, D.C., Hrapchak, B.B. Theory and Practice of Histotechnology, 2nd Edition. Battelle Press, Columbus, OH.

Storage: 2° C  25° C

**Mixed Storage Conditions.
Separate Contents.**

 ScyTek Laboratories, Inc.
 205 South 600 West
 Logan, UT 84321
 435-755-9848
 U.S.A.

 Authorized Representative in Europe
 (Regulatory affairs only)
 EmergoEurope (31)(0) 70 345-8570
 Molsnstraat 15
 2513 BH Hague, The Netherlands