

Hematoxylin, For Automation

Description: Produces a crisp blue nuclei. Hematoxylin, For Automation has been specifically formulated for situations in which a lighter blue nuclei is desired (IHC, etc.). This reagent is ideal for use with autostainers.

Uses/Limitations: Histology/Cytology.
Do not use past expiration date.

Availability:

<u>Item #</u>	<u>Volume</u>
HAQ500	500 ml
HAQ999	1000 ml
HAQ010	10 Liters

Storage: Room Temperature.
Keep container tightly closed.

Procedure:

1. Re-hydrate tissue as usual.
2. Apply enough Hematoxylin to completely cover specimen and incubate for 2-10 minutes with hematoxylin according to desired intensity.
3. Apply Bluing Reagent (ScyTek catalog# BRT500) for 5-10 seconds.
4. Rinse in DI water.
5. Coverslip using media of choice.

Precautions: Do not pipette by mouth.
Wash thoroughly after handling.
Reagent will stain clothing.