


Calcium Stain Kit (Modified Von Kossa)

Description: The Calcium Stain Kit (Modified Von Kossa) is intended for use in the histological visualization of calcium deposits in paraffin sections.

Calcium in mass deposits:	Black
Calcium in dispersed deposits:	Gray
Nuclei:	Red
Cytoplasm:	Light Pink

Uses/Limitations: Not to be taken internally.
For In-Vitro Diagnostic use only.
Histological applications.
Do not use if reagents become cloudy.
Do not use past expiration date.
Use caution when handling reagents.
Non-Sterile.

Control Tissue: TCS0007-5 (5 Slides)
TCS0007-25 (25 Slides)
Any paraffin embedded tissue that contains calcium deposits.




Ordering information regarding individual components on back page!


Kit Contents:

<u>Item #</u>	<u>Components</u>	<u>Volume</u>	<u>Storage</u>
SNV125	Silver Nitrate Solution (5%)	125 ml	2-8°C
STB125	Sodium Thiosulfate Solution (5%)	125 ml	18-25°C
NFS125	Nuclear Fast Red Solution	125 ml	18-25°C

Precautions: Avoid contact with skin and eyes.
Harmful if swallowed.
Follow all Federal, State, and local regulations regarding disposal.

Storage: 2° C  25° C

**Mixed Storage Conditions.
Separate Contents.**

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.

CE IVD

EC REP
Emergo Europe
Prinsessegracht 20
2514 AP The Hague, The Netherlands

Procedure (Standard):


1. Deparaffinize sections if necessary and hydrate to distilled water.
2. Incubate slide in Silver Nitrate Solution (5%) for 30-60 minutes while exposing to either ultraviolet light or incandescent light at 75 watts or greater. For best results, keep light source within 2 feet (61cm) of slide during Silver Nitrate staining procedure.
3. Rinse in 3 changes of distilled water.
4. Incubate slide in Sodium Thiosulfate Solution (5%) for 2 minutes.
5. Rinse for 2 minutes in running tap water followed by 2 changes of distilled water.
6. Stain tissue section with Nuclear Fast Red Solution for 5 minutes.
7. Rinse for 2 minutes in running tap water followed by 2 changes of distilled water.
8. Dehydrate very quickly in 3 changes of Absolute Alcohol.
9. Clear, and mount in synthetic resin.

References:


1. Sheenan, D.C., Hrapchak, B.B. Theory and Practice of Histotechnology, 2nd Edition. Battelle Press, Columbus, OH.
2. Clark, G., et al. Staining Procedures, 4th Edition, Williams & Wilkins Press, Baltimore, MD.
3. Symonds, D.A., Use of the Von Kossa stain in identifying occult calcifications in breast biopsies. American Journal of Clinical Pathology, 1990, July; 94(1) pages 44-48.

Bulk Reagent Ordering Information and Current Pricing at www.scytek.com

Description:	Catalog #	Volume
Silver Nitrate Solution (5%)	SNV125	125 ml
	SNV500	500 ml
Sodium Thiosulfate Solution (5%)	STB125	125 ml
	STB500	500 ml
	STB999	1000 ml
Nuclear Fast Red Solution	NFS125	125 ml
	NFS500	500 ml
	NFS999	1000 ml

 Storage: 2° C  25° C

**Mixed Storage Conditions.
Separate Contents.**



ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.


Emergo Europe
Prinsessegracht 20
2514 AP The Hague, The Netherlands