

Cytokeratin, Multi (Epithelial Marker); Clone C11 (Concentrate)

Availability/Contents:

<u>Item #</u>	<u>Volume</u>
RA0381-C.5	0.5 ml

Description:

Species: Mouse

Immunogen: Keratin-enriched preparation from cultured human A431

Clone: C11

Isotype: IgG1

Entrez Gene ID: 3851 (CK4); 3852 (CK5); 3853 (CK6A); 3854 (CK6B); 286887 (CK6C); 3856 (CK8); 3858 (CK10); 3860 (CK13); 3875 (CK18) (Human)

Hu Chromosome Loc.: 12q13.13 (CK4); 12q13.13 (CK5); 12q13.13 (CK6); 12q13.13 (CK8); 17q21.2 (CK10); 17q21.2 (CK13); 12q13.13 (CK18).

Synonyms: Multiple

Mol. Weight of Antigen: Multiple

Format: 200µg/ml of Ab purified from Bioreactor Concentrate by Protein A/G. Prepared in 10mM PBS with 0.05% BSA & 0.05% azide.

Specificity: This antibody recognizes cytokeratins 4, 5, 6, 8, 10, 13, and 18.

Background: This is a broad-spectrum antibody which has been reported to differentiate epithelial tumors from non-epithelial tumors. Many studies have shown the usefulness of keratins as markers in cancer research and tumor diagnosis.

Species Reactivity: Human, Cow, Rat, Mouse, Guinea pig, Frog, Goat, Marmoset, and Pig. Others not known.

Positive Control: A431 cells, skin, colon carcinoma.

Cellular Localization: Cytoplasmic

Titer/ Working Dilution: Immunohistochemistry (Frozen and Formalin-fixed): 0.5-1 µg/ml
Flow Cytometry: 0.5-1 µg/million cells
Immunofluorescence: 0.5-1 µg/ml
Western Blotting: 0.5-1 µg/ml

Microbiological State: This product is not sterile.

Storage: 2° C  8° C




ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.

CE

 EmergoEurope (31)(0) 70 345-8570
Molsnstraat 15
2513 BH Hague, The Netherlands

Uses/Limitations: Not to be taken internally.
 For Research Use Only.
 This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy.
 Do not use if reagent becomes cloudy.
 Do not use past expiration date.
 Non-Sterile.



Ordering Information and Current Pricing at www.scytek.com

Formalin-fixed, paraffin-embedded colon (10X) stained with Cytokeratin, Multi; Clone C11.

Procedure:


1. **Tissue Section Pretreatment (Required):** Staining of formalin fixed, paraffin embedded tissue sections is significantly enhanced by pretreatment with Citrate Plus (ScyTek catalog# CPL500).
2. **Primary Antibody Incubation Time:** We suggest an incubation period of 30 minutes at room temperature. However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
3. **Visualization:** For maximum staining intensity we recommend the “UltraTek HRP Anti-Polyvalent Lab Pack” (ScyTek catalog# UHP125, see IFU for instructions) combined with the “DAB Chromogen/Substrate Bulk Pack (High Contrast)” (ScyTek catalog# ACV500, see IFU for instructions).


Precautions: Contains Sodium Azide as a preservative (0.09% w/v).
 Do not pipette by mouth.
 Avoid contact of reagents and specimens with skin and mucous membranes.
 Avoid microbial contamination of reagents or increased nonspecific staining may occur.
 This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200, OSHA Hazardous Communication Standard and EC Directive 91/155/EC.

References:

1. Bartek J et. al. J Pathol, 1991, 164(3):215-24.
2. Lane EB; Alexander CM. Seminars in Cancer Biology, 1990, 1:165-79.
3. Bartkova J; et al. Neoplasma, 1991, 38:439-46.
4. Kasper M. Histochemistry, 1991, 95(6):613-20.


Warranty: No products or “Instructions For Use (IFU)” are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 2° C  8° C



ScyTek Laboratories, Inc.
 205 South 600 West
 Logan, UT 84321
 U.S.A.



 EmergoEurope (31)(0) 70 345-8570
 Molsnstraat 15
 2513 BH Hague, The Netherlands