

# Phosphotyrosine (P-Tyr); Clone PY793 (Concentrate)

**Availability/Contents:**
**Item #**

RA0420-C.5

**Volume**

0.5 ml

**Description:**

Species:	Mouse
Immunogen:	Phosphotyrosine conjugated to BSA
Clone:	PY793
Isotype:	IgG2b
Entrez Gene ID:	Not Applicable
Hu Chromosome Loc.:	Not Applicable
Synonyms:	P-Tyr
Mol. Weight of Antigen:	Depends upon the phosphorylated target
Format:	200µg/ml of Ab purified from Bioreactor Concentrate by Protein A/G. Prepared in 10mM PBS with 0.05% BSA & 0.05% azide.
Specificity:	This antibody shows no cross-reaction with other phosphoamino acids and is superb for multiple applications including staining of formalin-fixed, paraffin-embedded tissues.
Background:	Protein phosphorylation is a fundamental event in the regulation of a large number of intracellular processes. Phosphorylation of specific tyrosine residues is the result of activation or stimulation of their respective protein tyrosine kinases. The phosphorylated proteins can be auto-phosphorylated kinases or certain cellular protein substrates. Tyrosine-phosphorylated proteins are involved in signal transduction and in the regulation of cell proliferation. An antibody to phosphotyrosine provides an excellent tool for the detection, characterization, and purification of phosphotyrosine-containing proteins.
Species Reactivity:	All species.
Positive Control:	MCF-7, MDA-231, T47-D cells or breast carcinoma.
Cellular Localization:	Depends upon the location of phosphorylated target
Titer/ Working Dilution:	Immunohistochemistry (Frozen and Formalin-fixed): 1-2 µg/ml Flow Cytometry: 0.5-1 µg/million cells Immunofluorescence: 1-2 µg/ml Western Blotting: 0.5-1 µg/ml Immunoprecipitation: 1-2 µg/500µg protein lysate
Microbiological State:	This product is not sterile.

Storage: 2° C



8° C



ScyTek Laboratories, Inc.  
205 South 600 West  
Logan, UT 84321  
U.S.A.




EC REP

EmergoEurope (31)(0) 70  
345-8570  
Molsnstraat 15  
2513 BH Hague, The Netherlands

**Uses/Limitations:**

- Not to be taken internally.
- For Research Use Only.
- This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy.
- Do not use if reagent becomes cloudy.
- Do not use past expiration date.
- Non-Sterile.



Formalin-fixed, paraffin-embedded human breast carcinoma stained with Phosphotyrosine; Clone PY793.

**Ordering Information and Current Pricing at** [www.scytek.com](http://www.scytek.com)

### Procedure:

- Tissue Section Pretreatment (Highly Recommended):** Staining of formalin fixed, paraffin embedded tissue sections is significantly enhanced by pretreatment with Citrate Plus (ScyTek catalog# CPL500).
- Primary Antibody Incubation Time:** We suggest an incubation period of 30 minutes at room temperature. However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
- Visualization:** For maximum staining intensity we recommend the “UltraTek HRP Anti-Polyvalent Lab Pack” (ScyTek catalog# UHP125, see IFU for instructions) combined with the “DAB Chromogen/Substrate Bulk Pack (High Contrast)” (ScyTek catalog# ACV500, see IFU for instructions).

**Precautions:**


- Contains Sodium Azide as a preservative (0.09% w/v).
- Do not pipette by mouth.
- Avoid contact of reagents and specimens with skin and mucous membranes.
- Avoid microbial contamination of reagents or increased nonspecific staining may occur.
- This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200, OSHA Hazardous Communication Standard and EC Directive 91/155/EC.

### References:

- Hunter, T., et al. 1985. Protein-tyrosine kinases. Annu. Rev. Biochem. 54: 897-930.

**Warranty:**

No products or “Instructions For Use (IFU)” are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 2° C  8° C



ScyTek Laboratories, Inc.  
205 South 600 West  
Logan, UT 84321  
U.S.A.



EC REP

EmergoEurope (31)(0) 70  
345-8570  
Molsstraat 15  
2513 BH Hague, The Netherlands